

CLINICAL COMPARISON OF THYROGLOBULIN MEASUREMENT FOLLOWING THYROIDECTOMY AND RADIOIODINE ABLATION (TRA) USING TWO ULTRASENSITIVE METHODS FROM THE HILOTRIAL.

Turner, Steve¹; Neely, Dermot¹; Hackshaw, Allan²; Patel, Dhanesh³; Clark, Penny⁴; Franklyn, Jayne⁵; Harmer, Clive⁶; Yap, Beng⁷; Wadsley, Jonathan⁸; Clarke, Susan⁹; Moss, Laura¹⁰; Mallick, Ujjal¹¹

¹Newcastle upon Tyne NHS Foundation Trust, Newcastle upon Tyne, Tyne and Wear, United Kingdom (Great Britain); ²University College Hospital, London, United Kingdom (Great Britain); ³University College London, London, United Kingdom (Great Britain); ⁴University Hospitals Birmingham NHS Trust, Birmingham, UK, United Kingdom (Great Britain); ⁵University of Birmingham, Birmingham, United Kingdom (Great Britain); ⁶Royal Marsden Hospital, London, United Kingdom (Great Britain); ⁷Christie NHS Trust, Manchester, United Kingdom (Great Britain); ⁸Weston Park Hospital, Weston, United Kingdom (Great Britain); ⁹Guys and St Thomas' Hospital, London, United Kingdom (Great Britain); ¹⁰Velindre Cancer Centre, Cardiff, Wales, United Kingdom (Great Britain); ¹¹Department of Oncology, Freeman Hospital, Newcastle upon Tyne Hospitals NHS Trust, Newcastle upon Tyne, Tyne and Wear, United Kingdom (Great Britain)

Background/Purpose: Thyroglobulin (Tg) is used to detect recurrent thyroid cancer following TRA . Early detection improves outcome and has led to the development of sensitive assays. We aimed to review the clinical performance of two of these assays in the context of the HiLo trial.

Methods: From the HiLo trial 320 patients had Tg values measured on two assays under thyrogen stimulation 6-9 months following TRA. Measurement was performed using two second generation IRMA assays; Brahms assay and Beckman Coulter Access assay. Results were compared and assessed with clinical outcomes after a median follow up of 26 months.

Results: Detectable Tg without recurrence was found in 60 (Brahms, <0.5ug/L) and 164 (Beckman, <0.1ug/L) patients respectively. Single samples with concentrations of 0.5ug/L (Brahms) and 0.1ug/L (Beckman) measured on both assays (60 samples) showed good correlation (R²=0.988). Ninety seven patients were shown to have an undetectable level of Tg on the Brahms assay (<0.5ug/L) but detectable level using the Beckman assay (0.1-0.4ug/L). Seven patients had recurrence; 4 in thyroid bed who had positive Tg measurements using both assays, 2 with lymph node involvement did not have enough samples for Beckman assay analysis and one patient with lung secondary had undetectable Tg on both assays.

Discussion & Conclusion: Ultrasensitive assays provide additional benefit of detecting low serum Tg concentrations. So far during follow up no additional clinical benefit was demonstrated with an assay detection limit of 0.1ug/L in this randomised trial. Further analysis will be reported with longer follow up.